

## DECLARATION OF CONFORMITY

We declare under own responsibility that the following scales:

- *Non-approved Medical Scales*

**6163, 7815, 8310, 8320, 8352, 8777, 9202, 9203**

- *Electronic Height Rods*

**5003, 5002.01, 5002.02, 5001.10, 5005.01, 5005.10, 5005.11**

manufactured by us meet the provisions of the Council Directive

**2014/30/EU**

In general we meet the following directives:

Directive / Standard	
Company standard scale	FR 001 <sup>2</sup>
EMV directive	2014/30/EU
Standard for emission	EN60601-1-2
Standard for immunity	EN60601-1-2
immunity criterion	A (3V/m) <sup>4</sup>
ESD criterion	B
burst criterion	(B) <sup>3</sup>
Surge criterion	(B) <sup>3</sup>
Low voltage directive	2014/35/EU
Battery directive	2006/66/EC
REACH	(EG) No. 1907/2006
Food law	1935/2004/EC
Protection class	EN 60 529
RoHS	2011/65/EU

<sup>2)</sup> FR 001 = Factory Regulation 001

<sup>3)</sup> in the event that electrical wires are present external to the product

<sup>4)</sup> according to article 4 of the EMC act, a comment in German must be contained in the operating instructions, if the criterion cannot be maintained

**Please note:**

This document annuls the previous sent Declaration of Conformity for non-approved medical scales and electronic height rods.

**Validity:** until 31.12.2016

**Soehnle Industrial Solutions GmbH  
Gaildorfer Str. 6  
D-71522 Backnang, Germany**

**Michael Schläger  
General Manager**

21.07.2016