

DECLARATION OF CONFORMITY

We declare under own responsibility that our

- **Approved Medical Scales**

7702.37.002, 7702.65.002, 7707.37.002, 7708.37.002, 7708.65.001, 7708.65.002, 7708.37.502, 7708.65.501, 7708.65.502, 7708.86, 7708.87, 7708.88, 7710.01.001, 7710.02.001, 7711.01.001, 7711.02.001, 7725.01.001, 7730.01.001, 7731.01.001, 7732.01.001, 7732.30.001, 7732.10.001, 7771.37.002, 7771.65.002

- **Indicators**

3020.05.001, 3020.05.002, 3705.15.001, 3705.55.001, 3710.15.001, 3710.55.001

manufactured by us meet the provisions of the
Council Directive **2009/23/EC** and MDD **93/42/EEC (2007/47/EC)**

In general we meet the following directives:

Directive / Standard Classification	
Non automatic weighing instrument directive	2009/23/EC Class III
Scale standard	EN45501
OIML ¹	R76
EMV directive	2004/108/EC
Standard for emission	EN60601-1-2
Standard for immunity	EN45501/60601-1-2
immunity criterion	A
ESD criterion	A
burst criterion	A
Surge criterion	A
Low voltage directive	2006/95/EC
Standard for electr. Safety	EN60601-1
Battery directive	2006/66/EC
REACH directive	EC 1907/2006
Medical device directive	93/42/EEC (2007/47/EC) Class I m
Food law	1935/2004/EC
Protection class	EN 60 529
RoHS	2011/65/EU

¹⁾ The decision to make an application for an OIML certificate should be made, in each case, when an application is made for a type approval

Please note:

This document annuls the previous sent Declaration of Conformity for approved medical scales and indicators.

Validity: until 31.12.2016

Soehnle Industrial Solutions GmbH
Gaidorfer Strasse 6
D-71522 Backnang, Germany

Michael Schläger
General Manager
05.05.2015